

# Prepare for Your Upcoming Multi-Level Guided Surgery

Provided by Glidewell

*See Reverse for Required Instruments and Materials*



## Tier 1

- 3D-Printed Model
- Anchor Pin Mounting Guide
- Foundation (Bone Reduction) Guide
- Osteotomy Guide

## Tier 2

- 3D-Printed Bone Reduction Model
- Anchor Pin Mounting Guide
- Foundation (Bone Reduction) Guide
- Osteotomy Guide
- Prosthetic Delivery Guide
- Four Multi-Unit Abutments
- Four Stock Temporary Cylinders\*
- Immediate Fixed Screw-Retained Provisional Implant Prosthesis

## Tier 3

- 3D-Printed Bone Reduction and Analog Model
- Anchor Pin Mounting Guide
- Foundation (Bone Reduction) Guide
- Anchor Pins
- Osteotomy Guide
- Prosthetic Delivery Guide
- Multi-Unit Abutments
- Multi-Unit Temporary Healing Caps
- Custom Gasket for Relining Prostheses
- Stock Temporary Cylinders\*
- Implant Verification Prosthesis
- Custom Precut Temporary Cylinders\*
- Immediate Fixed Screw-Retained Provisional Implant Prosthesis
- Blockout Shims

### Additional Items (Fees Apply):

- Hahn™ Tapered Implants
- Additional Temporary Cylinders (Tier 2 Only)
- Additional Multi-Unit Abutments (Tier 2 Only)
- Anchor Pins (Tiers 1 and 2 Only)
- Anchor Pin Drills (Tiers 1, 2 and 3)
- Multi-Unit Temporary Healing Caps (Tiers 1 and 2 Only)

\*For flat-rate fee cases, Inclusive® Bite Verification Cylinders are provided. For other cases, original equipment (OEM) temporary cylinders are provided and incur an additional fee.

866-497-3692

[glidewell.com/dtp](http://glidewell.com/dtp)

 **Glidewell**  
Implant Solutions

## Surgical Instruments and Components Needed (Applies to Tiers 1, 2 and 3)

- Implants (verify type, diameter and length on surgical report)
- Anchor Pins (verify quantity and length on surgical report)
  - 13 mm
  - 15 mm
- 1.5 mm Anchor Pin Drill (available through Glidewell DTP by request)
- Bone Reduction Burs
  - Alveoplasty Carbide Burs (straight or tapered recommended)
- Handpiece
  - 1:1 for bone contouring
- Desired Implant Guided Surgery Kit

## Restorative Instruments and Materials Needed for Delivery of Provisional Prosthesis (Applies to Tiers 2 and 3)

- Lab Handpiece / Motor
- Suggested Burs (make sure the burs fit handpiece)
  - Round #8 for adjusting the temporary cylinders
  - Football-shaped crosscut bur for adjusting and smoothing
  - 1557 carbide bur for refining
  - Round carbide bur, #1 or #2, for refining and opening air bubbles in acrylic
  - Rubber polishing points
- Luting / Bonding Material for picking up temporary cylinders  
(Voco® Quick Up, and Stellar™ pick-up recommended)
- Teflon® tape (to protect the prosthetic screw in the access channels)
- Composite or temporary filling material to seal access hole
- Compressed air for cleaning debris during finishing of the provisional  
(if air supply is not available, a can of compressed air can be utilized)
- Tissue conditioner or soft-reline material to be utilized in the back-up denture in case the long-term provisional cannot be delivered on the day of surgery
- Articulating paper for occlusal verification and adjustments
  - Ideally:
    - Minimal working and non-working contacts
    - Evenly distributed centric contacts
- Light Cure Unit
  - Handheld unit is recommended for both intraoral and extraoral curing